

**CALIBRATION CERTIFICATE**

ULR No : CC269523000000352	Certificate No : BFI/Bangalore /240814/NABL/11/5	Page Number : 01/01
----------------------------	--	---------------------

Customer Name & Address	Labfix Durban	Date Recipet	
		Date Calibration	14/08/2024
		Expire Date	13/08/2025
		Date Issue	14/08/2024

**DETAILS OF UNIT UNDER Testing (UUT)**

Name of Instrument	: Oven	Inward No.	: LabFix/240814/11
Manufacturer	: BFI	Identification No.	: Labfix004
Model	: bfi	Serial no.	: 123
Range	: 0 to 150 °c	Accuracy	:
Resolution/Least Count	: 0.01 °c	Location	: Main Line
Calibration Location	: Onsite	Condition of Instrument	: Ok
Calibration Procedure	: BFI102	Standard Reference	:

Environmental Conditions:	Temperature : 25.06 °C	Humidity: 55.8 %Rh
---------------------------	------------------------	--------------------

**Details of Calibration Test Equipment Used**

UUT Name	Id/Sr. No.	Certificate No.	Calibrated By	Calibration Expire Date
4 wired RTD Sensor With Scanner	GC-054 to 069	GC/AV/P0445B/07/24	GC Pune	14/08/2025

DISCIPLINE : Thermal - Temperature

**Calibration Results**

Set Value	Indicated Value on UUT	Sensor Id	Test Readings Mean	Deviation	Expanded Uncertainty (±)
°c	°c		°c	°c	°c
0	0.00	1	0.003	-0.003	0.64
10	10.00	2	10.004	-0.004	0.64

Remarks : Ladder of Success



Print At : 14/08/2024, 13:15

For BFI, Mumbai		For Customer
Calibrated By	Checked And Approved By	Checked By
Electronically signed by bfi 14/08/2024, 13:09	Electronically signed by bfi 14/08/2024, 13:15	
Calibrated By	Technical Signatory	Labfix

----- END OF CALIBRATION CERTIFICATE -----