



PRECISION INSTRUMENTATION AND SERVICES PVT.LTD.

ISO/IEC 17025:2017 Accredited Calibration Laboratory

◆ Instrument Calibration ◆ AHU Validation ◆ Thermal Validation ◆ PLC Validation

21. Shreeram Bhavan, F-9, Shreeramkunj, Takli Road, Dwarka, Nashik - 422 011.

M.: 9860102524. Website: www.pisnashik.com E-mail: sales1@pisnashik.com

CALIBRATION CERTIFICATE

Certificate No : PISPL/SAMB/250220/2067/1

Page Number : 01/01

1.Customer

Ajanta Pharma Ltd.

Gut No.11/12/14/15, Chitegaon,Paithan Road, Chhatrapati Sambhajnagar, Maharashtra

Date of Receipt : 20/02/2025
Date of Calibration : 20/02/2025
Due Date Calibration : 19/08/2025
Date of Issue : 20/02/2025
Service Request No : SRF/SAMB/250220/2067
Calibration Method : PISPL/ET/WI/08
Condition of UUC : Satisfactory
Location of Calibration : Onsite

2.Environment Condition:

Ambient Temp:22.9 °C

Relative Humidity: 51.4 %RH

3.Description of UUC

Instrument Name : Timer Range : 9 Hrs, 59 Min, 59 Sec
Instrument ID No : N/PR/FTA/002/TM-01 Resolution : NA
Make / Model No : Electrolab Specified Accuracy : ± 2 % of set value
Type : Digital Location : Friability Tester
Operating Range : 0 to 4 min

4.Details of Standard Used For Calibration

Name	Make	ID No./Sr.No.	Certificate No.	Certified By	Calibration Validity
Digital Stop Watch	Racer	PISPL/ABAD/STW/01	E2024/12/1697	PISPL	04/01/2026

5.Calibration Results

Calibration Point	UUC Reading	Standard Reading	Error
min	min	mm:ss:s/100	mm:ss:s/100
4	4	4:00:06	-0:00:06

Remarks :

The results comply with acceptance criteria

- Note:
- 1) UUC stands for Unit Under Calibration.
 - 2) This certificate refers only to the particular item submitted for calibration
 - 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Precision Instrumentation And Services Pvt. Ltd., Nashik".
 - 4) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

For Precision Instrumentation And Services Pvt.Ltd.

For Customer

Calibrated By	Checked By	Checked By	Approved By
Electronically signed by Mr.Ravindra Tattu 20/02/2025, 11:05	Electronically signed by Ms.Akanksha Salve 21/02/2025, 11:14		
Calibration Engineer	Senior Engineer	Ajanta Pharma Ltd.	Ajanta Pharma Ltd.

----- END OF CALIBRATION CERTIFICATE -----