



WORLD ONE INSTRUMENTS PVT.LTD.

Calibration Laboratory



CC-4040

ISO/IEC 17025 ACCREDITED CALIBRATION LABORATORY BY NABL VIDE CERTIFICATE NUMBER CC-4040

#47, Ganapathi Nagar, Peenya Industrial Area, Bengaluru - 560 058

Ph. 99000 99913 / 99000 99912 / 99000 32111 | Email: worldoneinstruments@gmail.com | info@wois.in | Web: www.wois.in

CALIBRATION CERTIFICATE

ULR Number: CC37682400000146F

Certificate Number: R1/2503/C/2/81

SRF No	R1/2503/C/2	Date of Issue	11/04/2025
		Date Calibration	03/04/2025
		Calibration Due on	02/04/2026

Name & Address of the Customer	Worldone instruments Pvt Ltd 47,Ganapathy Nagar , Peenya Industrial area,Bengaluru-560058	Status of the item on Receipt	Satisfactory
		Date of Receipt	20/03/2025
		Calibration Location	In House
		Usage Location	Lab

Description & Identification of Item (DUC)

Nomenclature	: Micropipette	Make	: Thermo scientific
ID No	: WOIS-MP-01	Range / Capacity	: 100 to 1000 µL
Resolution	: 0.01 µL		

Environmental Conditions:	Temperature: 23.4 °C	Relative Humidity: 50 %RH
---------------------------	----------------------	---------------------------

Calibration Procedure:	WOI-ME-SOP-VOLUME	Calibration Method:
------------------------	-------------------	---------------------

Calibrated By : QM

Reference Equipment used for Calibration

Nomenclature	Serial No.	Certificate No	Validity
Digital Weighing Balance	3378593	2024-25/EQN/24080003-0001	31/07/2025

Discipline : Mechanical (Group : Volume)

Calibration Results

Nominal Value On UUC	Average Value on Standard	Error	Expanded Uncertainty (±)
µL	µL	µL	µL
100.00	99.374321	0.625679	0.48
500.00	501.690487	-1.690487	0.8187272
1000.00	993.301751	6.698249	1.6207989

Remarks:

The results comply with acceptance criteria.



Checked By	Quality Manager	Authorized Signatory
		Electronically Signed By QM 11/04/2025, 15:06
		Quality Manager

This certificate is electronically signed and does not required physical signature

- End of Calibration Certificate -