

**CALIBRATION CERTIFICATE**

<b>ULR Number:</b> CC26952300000055F	<b>Certificate Number:</b> BFI/Maharashtra/20250427/NABL/13/4
--------------------------------------	---

<b>Customer Name &amp; Address</b>	Well Done Labs Madurai	<b>Date of Receipt</b>	27/04/2025
		<b>Date Calibration</b>	01/05/2025
		<b>Due Date</b>	30/04/2026
		<b>Date Issue</b>	15/05/2025

**DETAILS OF UNIT UNDER CALIBRATION (UUC)**

<b>Name of Instrument</b>	: Pressure Indicator	<b>SRF No</b>	: SRF/Maharashtra/250427/NABL/13
<b>Make</b>	: Wika	<b>Identification No</b>	: WD-04
<b>Range</b>	: 0 to 40 bar	<b>Resolution/Least Count</b>	: 0.1 bar
<b>Location</b>	: -	<b>Calibration Location</b>	: Onsite
<b>Condition of Instrument</b>	: Satisfactory	<b>Calibration Procedure</b>	: TSC/CAL/601

<b>Environmental Conditions:</b>	<b>Temperature:</b> 22 °C	<b>Humidity:</b> 22 %RH
----------------------------------	---------------------------	-------------------------

**DETAILS OF STANDARD USED FOR CALIBRATION**

Instrument Name	Id/Sr. No.	Certificate No	Calibrated By	Calibration Due Date
Digital Pressure Gauge	Bfi003	C&IJ/CAL/23-05/074	Transcal	18/02/2026

**Discipline :** Mechanical - Pressure Indicating devices

**Calibration Results**

Set pressure on UUC	Reading on Master Equipment			Mean Value	Error	Repeatability	Hysteresis	Expanded Uncertainty (±)	Compliance Status
	m1	m2	m3	Miw ((M1+M3)/2 +M2)/2	Error	(M3-M1)	(M2-M1)		
bar	bar	bar	bar	bar	bar	bar	bar	bar	
0.0	0.00001	0.00003	0.00003	0.00002	-0.00002	0.00002	0.00002	0.11	Pass
5.0	2.00001	2.00001	2.00001	2.00001	2.99999	0.00000	0.00000	0.11	Pass
10.0	5.00002	5.00002	5.00002	5.00002	4.99998	0.00000	0.00000	0.11	Pass



For BFI, Mumbai		For Customer
<b>Calibrated By</b>	<b>Checked And Approved By</b>	<b>Reviewed By</b>
Pritam	bfi	
	<i>Bhushan</i>	
<b>Calibration Engineer</b>	<b>Quality Manager</b>	<b>Well Done Labs</b>