

CALIBRATION CERTIFICATE

ULR Number: CC269523000000055F	Certificate Number: BFI/Maharashtra/20250427/NABL/13/4
--------------------------------	--

Customer Name & Address	Well Done Labs Madurai	Date of Receipt	27/04/2025
		Date Calibration	01/05/2025
		Due Date	30/04/2026
		Date Issue	15/05/2025

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Name of Instrument	: Pressure Indicator	SRF No	: SRF/Maharashtra/250427/NABL/13
Make	: Wika	Identification No	: WD-04
Range	: 0 to 40 bar	Resolution/Least Count	: 0.1 bar
Location	: -	Calibration Location	: Onsite
Condition of Instrument	: Satisfactory	Calibration Procedure	: TSC/CAL/601
Sensor Type	: E-Type		

Environmental Conditions:	Temperature: 22 °C	Humidity: 22 %RH
---------------------------	--------------------	------------------

DETAILS OF STANDARD USED FOR CALIBRATION

Instrument Name	Id/Sr. No.	Certificate No	Calibrated By	Calibration Due Date
Digital Pressure Gauge	Bfi003	C&IJ/CAL/23-05/074	Transcal	18/02/2026

Discipline : Mechanical - Pressure Indicating devices

Calibration Results

Set pressure on UUC	Reading on Master Equipment			Mean Value	Error	Repeatability	Hysteresis	Expanded Uncertainty (±)	Compliance Status
	m1	m2	m3	Miw ((M1+M3)/2 +M2)/2	Error	(M3-M1)	(M2-M1)		
bar	bar	bar	bar	bar	bar	bar	bar	bar	
0.0	0.00001	0.00003	0.00003	0.00002	-0.00002	0.00002	0.00002	0.11	Pass
5.0	2.00001	2.00001	2.00001	2.00001	2.99999	0.00000	0.00000	0.11	Pass
10.0	5.00002	5.00002	5.00002	5.00002	4.99998	0.00000	0.00000	0.11	Pass



For BFI, Mumbai		For Customer
Calibrated By	Checked And Approved By	Reviewed By
Electronically Signed By Pritam 01/05/2025, 10:14	Electronically Signed By bfi 15/05/2025, 15:37	
Calibration Engineer	Quality Manager	Well Done Labs

This certificate is electronically signed and does not require physical signature

- End of Calibration Certificate -

Page No : 01 of 01