

CALIBRATION CERTIFICATE

ULR Number: CC26952300000055F	Certificate Number: BFI/Maharashtra/20250427/NABL/13/4
--------------------------------------	---

Customer Name & Address	Well Done Labs Madurai	Date of Receipt	27/04/2025
		Date Calibration	01/05/2025
		Due Date	30/04/2026
		Date Issue	15/05/2025

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Name of Instrument	: Pressure Indicator	SRF No	: SRF/Maharashtra/250427/NABL/13
Make	: Wika	Identification No	: WD-04
Range	: 0 to 40 bar	Resolution/Least Count	: 0.1 bar
Location	: -	Calibration Location	: Onsite
Condition of Instrument	: Satisfactory	Calibration Procedure	: TSC/CAL/601
Sensor Type	: E-Type		

Environmental Conditions:	Temperature: 22 °C	Humidity: 22 %RH
----------------------------------	---------------------------	-------------------------

DETAILS OF STANDARD USED FOR CALIBRATION

Instrument Name	Id/Sr. No.	Certificate No	Calibrated By	Calibration Due Date
Digital Pressure Gauge	Bfi003	C&IJ/CAL/23-05/074	Transcal	18/02/2026

Discipline : Mechanical - Pressure Indicating devices
Calibration Results

Set pressure on UUC	Reading on Master Equipment			Mean Value	Error	Repeatability	Hysteresis	Expanded Uncertainty (±)	Compliance Status
	m1	m2	m3	Miw ((M1+M3)/2 +M2)/2	Error	(M3-M1)	(M2-M1)		
bar	bar	bar	bar	bar	bar	bar	bar	bar	
0.0	0.00001	0.00003	0.00003	0.00002	-0.00002	0.00002	0.00002	0.11	Pass
5.0	2.00001	2.00001	2.00001	2.00001	2.99999	0.00000	0.00000	0.11	Pass
10.0	5.00002	5.00002	5.00002	5.00002	4.99998	0.00000	0.00000	0.11	Pass

Amendment Certificate No: BFI/Maharashtra/20250427/NABL/13/4

* This is the Amendment certificate of the BFI/Maharashtra/20250427/NABL/13/4 dated 01/05/2025, issued on 15/05/2025.

* Reason for Amendment:



For BFI, Mumbai		For Customer
Calibrated By	Checked And Approved By	Reviewed By
Electronically Signed By Pritam 01/05/2025, 10:14	Electronically Signed By bfi 15/05/2025, 15:37	
Calibration Engineer	Quality Manager	Well Done Labs

This certificate is electronically signed and does not required physical signature
 - End of Calibration Certificate -